

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3007
CALIBRATION DATE: 29-Mar-16

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.07210683e+001
h = 1.53765820e+000
i = 3.72566721e-004
j = 4.92239080e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.63938	0.00000	0.00000
-1.0000	34.8016	2.80350	5.01535	2.80350	-0.00000
1.0000	34.8020	2.97486	5.12488	2.97486	0.00000
15.0000	34.8018	4.27002	5.88689	4.27002	-0.00000
18.5000	34.8010	4.61656	6.07450	4.61656	0.00000
29.0001	34.7985	5.69974	6.62656	5.69974	-0.00000
32.5000	34.7880	6.07161	6.80570	6.07161	0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

